

UL STAFF STATION

APPLICATION

The IR-310E Staff Station provides reliable two-way, hands-free communication between the nurse master station and utility room, examination rooms, kitchen, or any location where staff members may need to originate calls or where cord sets are not required.

Calls are placed by use of a momentary push button. Solid-state lights on the panel indicate that the call has been placed and that the intercom is in use.

As an additional feature, the staff station is field programmable to function as a patient station, if desired.

FEATURES

- UL® 1069 Listed
- Supervised • Microprocessor design
- Solid-state multiplex circuitry
- Field programmable for staff or patient station operation
- Programmable call priority • All stations interchangeable
- Simplified multiplex wiring using plug-in type connectors
- Attractive flame retardant (UL® 94V-0) plastic panel

SPECIFICATIONS

Dimensions: Height: 4.5" (119 mm)
Width: 6.375" (162 mm)
Depth: 1.875" (48 mm)
Projects: 0.5" (12.7 mm)
above mounting surface

Construction: Flame-retardant (UL® 94V-0) plastic

Finish: Light texture, white

Speaker: 2.5" × 2.5" (64 mm × 64 mm)

Connections: Plug-in type

Housing: Steel City #H3BD with Steel City #3GC plaster ring or exact equal. Horizontal mount.

Wiring: Multiplex

REQUIRED COMPONENTS

Tek-CARE® NC-300II Nurse Call System



IR-310E
Staff Station

ARCHITECTS' AND ENGINEERS' SPECIFICATIONS

Staff station shall be TekTone® IR-310E, or approved equal as supplied by LEE DAN®. Model IR-310E shall include solid-state LSI/microprocessor technology, combination yellow callplaced indicator and call button, and combination red inuse indicator and reset button.

The station shall utilize a 2.5" × 2.5" speaker/microphone for accurate voice reproduction. The station shall be field programmable to emulate a patient station, if so desired.

All station components shall be mounted on an attractive, flame retardant (UL® 94V-0) plastic panel. Connections shall be plug-in type.

The IR-310E shall be UL® 1069 Listed.